2025/11/04 16:21

## **OSMOSERS-TROUBLESHOOTING**

PROBLEMS	POSSIBLES CAUSES	SUGGESTED ACTION		
- No production of purified water		indicator <i>PUMP</i> switched off	- Power switch is on OFF position.	Switch ON the device
		- The water supply of the tap closed or insufficient pressure (in that case the light "SOURCE" is on).	Check the water supply	
		- water hose folded or obstructed	Check that the inlet pipe is not damaged.	
<ul> <li>No production of purified water - Power light ON - all control light indicators OFF</li> </ul>	- on power supply	Check and replace the fuses		
- Not or low purified water production - indicator <i>PUMP</i> light ON - Pump running	Pre-treatment clogged	Replace the pre-treatment filters before the RO membrane		
		Increase the flow of the inlet water		
	Reverse osmosis membrane clogged	Change the membrane		
<ul> <li>Not or low purified water production - indicator PUMP light ON - Pump always OFF</li> </ul>	Pump fuse burned	Replace the pump fuse		
Light indicators FULL and PUMP alternate	Defectiveness of the high-pressure sensor or check valve.	Pressure sensor and check valve		
Lack of pressure and low flow output	- Watertank hand valve closed.	Turn ON the watertank hand valve		
	<ul> <li>volume of pure water need above the specification of the water purifier</li> </ul>	Wait the watertank is filled again		
	- Lack of pressure in the pressurised watertank	Re-adjust the air-pressure of the water tank		
purified water conductivity too high	Resin saturated.	Replace resin cartridge or bottle		
- POWER & FULL indicators ON - Pump OFF - Drain flowing continuously	defective	Replace the inlet solenoid valve		
Conductivity meter display OFF	Batteries out of order	Replace the batteries		

From:

https://diasys-technologies.com/waterpurifiers/ - DiaSys Water Purifier Systems

Permanent link:

× https://diasys-technologies.com/waterpurifiers/doku.php?id=rotroubleshooting&rev=1554383030

Last update: 2019/04/04 13:03