
2025/07/01 18:15 1/2 Introduction

Introduction

The water purifiers proposed by DIASYS are a combination of 2 elements:



The reverse osmosis membrane has the advantage to prolong the lifetime of the deionising filter.

The water purifiers designed by Diasys are especially built for providing deionised water to clinical chemistry and immunology analysers.

catalog

- Water purifiers
- UV sterilizer
- Auto-filling water level sensor
- Conductivity meters

recommended conditions

A reverse osmosis system works well with water at ambiant temperature. The tap water should not be too hot (above 35°C) or too cold (below 10°C). Outside these bounds, the membrane might be damaged irreversibly. The optimum production of water is obtained at 25°C.

In order to activate the reverse osmosis process, the tap water pressure in the membrane must be strong enough, above 2.5 BAR (ideal pressure is generally around 4.5 BAR) but can vary depending of the membrane. When the tap water pressure is too low or vary, a booster pump is required. It is recommended to rince the membrane every month using the following process:

- disconnect the restrictor
- leave water flowing during 20 minutes
- reconnect the restrictor

This manual process is similar to the flush function available on the system equipped with an electronic controller, which allows to rince the membrane easily just by pressing a button.

Q & A

Last update: 2019/04/04 10:24

- What's the difference between RO and DI water purification?
- Why carbon to filter water?
- Why a booster pump?
- How to disconnect autofilling? (OMINI autofilling version only)
- How to replace an old controller without white wire by a new one?
- Why indicator lamps of the controller are off?
- When to replace the reverse osmosis membrane?
- Is the UV sterilizer to be connected before or after the water purifier?

Troubleshooting

troubleshooting

From:

https://diasys-technologies.com/waterpurifiers/ - DiaSys Water Purifier Systems

Permanent link:

https://diasys-technologies.com/waterpurifiers/doku.php?id=start&rev=1554373475

Last update: 2019/04/04 10:24

×