

Membrane replacement test

- Measure the conductivity after the reverse osmosis membrane
- Measure the conductivity of the tap water
- Do the following calculation : $100 * (\text{conductivity reverse osmosis}) / (\text{conductivity raw water})$
- If the value is less **than 10%**, the membrane is still usable

In the other case, the reverse osmosis membrane begins to saturate.

From:

<http://diasys-technologies.com/waterpurifiers/> - **DiaSys Water Purifier Systems**

Permanent link:

<http://diasys-technologies.com/waterpurifiers/doku.php?id=whenreplacingmembrane>

Last update: **2019/03/27 10:12**

